Certificate of Analysis

Amended

Compliance | Adult Use

SC-000018,AU-SC-000106
ACT Laboratories (MI)
617 E. Hazel, Lansing, Michigan
5178621857
blakefeighner@actlab.com

Nowadays

Order No.: OMIELBN0903-0008682
150 Progress
California, 92618
info@trynowadays.com
7143674386

Sample: SMIELBN0915-IBEV-0006157

Unit Weight: 45.0000g
Sample Size: 2.0000 Units
Date Sampled: 09/03/2025 13:25
Sample Received: 09/03/2025 13:33
Report Created: 09/30/2025 12:46
Sampling SOP 204

CR-NOW750-05-32 - Nowadays - Original - 5mg

Ingestible, Beverage
Batch ID: CR-NOW750-05-32, Lot ID: See Page 9 for more details

Results

5.1277 mg/unit D9-THC	**0.1680 mg/unit** CBG
0.0000 mg/unit Total CBD	**5.1277 mg/unit** Total THC
5.2957 mg/unit Total Cannabinoids	**0%** Total Terpenes

Tests Summary

Cannabinoids Tested	**Moisture** Not Tested	**Microbials** Pass
Water Activity Not Tested	**Foreign Matter** Pass	**Terpenes** Tested
Phytol Tested	**Residual Solvents** Pass	**VEA / MCT Oil** Not Tested
Mycotoxins Pass	**Heavy Metals** Pass	**Pesticides** Pass
Miscellaneous Tested		

Dominant Terpenes

No terpenes detected

ACCREDITED CERT #7425.03

* Note: Analytes tested at ACT Laboratories of FL (ISO/IEC 17025:2017 #7425.05): 1,1-Dichloroethene; Captan, Chlordane, Chlormequat Chloride, Coumaphos, Dimethomorph, Mevinphos, PCNB, Piperonyl Butoxide, Spinetoram, Fenhaxamide.

Brad Cavinder
Laboratory Manager

♦ indicates ISO/IEC 17025:2017 accreditation is pending. This sample has been tested by ACT Laboratories using valid, ISO/IEC 17025:2017 accredited testing methodologies and a quality system as required by state law. Values reported relate only to the sample tested. Results are deemed acceptable if they fall within the specified acceptance criteria as defined by the applicable regulatory limit or client specification, with no measurement uncertainty applied. ACT Laboratories makes no claims as to the efficacy, safety or other risks associated with any detected or non-detected levels of any compounds reported herein. This Certificate shall not be reproduced except in full, without the written approval of ACT Laboratories. The authenticity of this document is only guaranteed if issued from an @actlab.com email address.

Certificate of Analysis

Amended

Compliance Adult Use

SC-000018,AU-SC-000106
ACT Laboratories (MI)
617 E. Hazel, Lansing, Michigan
5178621857
blakefeighner@actlab.com

Nowadays
Order No.: OMIELBN0903-0008682
150 Progress
California, 92618
info@trynowadays.com
7143674386

Sample: SMIELBN0915-IBEV-0006157
Unit Weight: 45.0000g
Sample Size: 2.0000 Units
Date Sampled: 09/03/2025 13:25
Sample Received: 09/03/2025 13:33
Report Created: 09/30/2025 12:46
Sampling SOP 204

CR-NOW750-05-32 - Nowadays - Original - 5mg
Ingestible, Beverage
Batch ID: CR-NOW750-05-32, Lot ID: See Page 9 for more details

Cannabinoids

Tested

SOP: SOP 801, SOP 301
Instrument: HPLC 1220 Agilent - 1325

Analyte	LOQ (%)	%	mg/g	mg/unit
CBDVa	0.000	ND	ND	ND
CBDV	0.000	ND	ND	ND
CBDa	0.000	ND	ND	ND
CBGa	0.000	ND	ND	ND
CBG	0.000	**0.0004**	**0.0037**	**0.1680**
CBD	0.000	< LOQ	< LOQ	< LOQ
THCV	0.000	< LOQ	< LOQ	< LOQ
THCVa	0.000	ND	ND	ND
CBN	0.000	< LOQ	< LOQ	< LOQ
D9-THC	0.000	**0.0114**	**0.1139**	**5.1277**
D8-THC	0.000	ND	ND	ND
CBC	0.000	< LOQ	< LOQ	< LOQ
THCa	0.000	< LOQ	< LOQ	< LOQ
CBCa	0.000	ND	ND	ND
Total CBD		0.0000	0.0000	0.0000
Total THC		**0.0114**	**0.1139**	**5.1277**
Total Cannabinoids		**0.0118**	**0.1177**	**5.2957**

Report notes:
Total THC = THCa * 0.877 + Δ9-THC; Total CBD = CBDa * 0.877 + CBD. LOQ = Limit of Quantitation, ND= Non-Detect; Cannabinoid potency values are reported by percentage of dry weight determined via loss on drying; Unless otherwise stated all quality control samples performed within specification established by the Laboratory.

Microbials

Pass

SOP: SOP 401-MI, SOP 425, SOP 418-MI

Analyte	Limit (CFU/g)	CFU/g	Status
Coliforms	100	ND	Passed
E. Coli	0	ND	Passed
Salmonella	0	ND	Passed
Aspergillus Flavus	0	ND	Passed
Aspergillus Fumigatus	0	ND	Passed
Aspergillus Niger	0	ND	Passed
Aspergillus Terreus	0	ND	Passed
Aspergillus Species	0	ND	Passed
STEC E. Coli	0	ND	Passed
Yeast & Mold QL	10,000	ND	Passed

Report notes:
TNTC = Too Numerous to Count, ND= Non-Detect; Unless otherwise stated all quality control samples performed within specification established by the Laboratory

Foreign Matter

Pass

SOP: SOP 203-MI
Date/Time Tested: 09/04/2025 12:45

Analyte	Limit (%)	%	Status
FM Stems	5.00	0.00	Passed
FM Organic	2.00	0.00	Passed
FM Inorganic		Absent	Passed

ACCREDITED
CERT #7425.03

* Note: Analytes tested at ACT Laboratories of FL (ISO/IEC 17025:2017 #7425.05): 1,1-Dichloroethene; Captan, Chlordane, Chlormequat Chloride, Coumaphos, Dimethomorph, Mevinphos, PCNB, Piperonyl Butoxide, Spinetoram, Fenhaxamide.

Brad Cavinder
Laboratory Manager

♦ indicates ISO/IEC 17025:2017 accreditation is pending. This sample has been tested by ACT Laboratories using valid, ISO/IEC 17025:2017 accredited testing methodologies and a quality system as required by state law. Values reported relate only to the sample tested. Results are deemed acceptable if they fall within the specified acceptance criteria as defined by the applicable regulatory limit or client specification, with no measurement uncertainty applied. ACT Laboratories makes no claims as to the efficacy, safety or other risks associated with any detected or non-detected levels of any compounds reported herein. This Certificate shall not be reproduced except in full, without the written approval of ACT Laboratories. The authenticity of this document is only guaranteed if issued from an @actlab.com email address.

Certificate of Analysis

Amended

Compliance Adult Use

SC-000018,AU-SC-000106
ACT Laboratories (MI)
617 E. Hazel, Lansing, Michigan
5178621857
blakefeighner@actlab.com

Nowadays
Order No.: OMIELBN0903-0008682
150 Progress
California, 92618
info@trynowadays.com
7143674386

Sample: SMIELBN0915-IBEV-0006157
Unit Weight: 45.0000g
Sample Size: 2.0000 Units
Date Sampled: 09/03/2025 13:25
Sample Received: 09/03/2025 13:33
Report Created: 09/30/2025 12:46
Sampling SOP 204

CR-NOW750-05-32 - Nowadays - Original - 5mg
Ingestible, Beverage
Batch ID: CR-NOW750-05-32, Lot ID: See Page 9 for more details

Terpenes

Tested

SOP: SOP 603-MI
Instrument: GCFID 8890 Agilent - 2603

Analyte	LOQ (%)	mg/g	%
a-Pinene	0.0084	ND	ND
Camphene	0.0084	ND	ND
Sabinene	0.0084	ND	ND
b-Pinene	0.0084	ND	ND
b-Myrcene	0.0084	ND	ND
a-Phellandrene	0.0084	ND	ND
d-3-Carene	0.0084	ND	ND
a-Terpinene	0.0084	ND	ND
Limonene	0.0084	ND	ND
p-Cymene	0.0084	ND	ND
Eucalyptol	0.0084	ND	ND
g-Terpinene	0.0084	ND	ND
Terpinolene	0.0084	ND	ND
Fenchol	0.0084	ND	ND
Sabinene Hydrate	0.0167	ND	ND
Linalool	0.0084	ND	ND
Fenchone	0.0084	ND	ND
Isopulegol	0.0084	ND	ND
Camphor	0.0084	ND	ND
Isoborneol	0.0084	ND	ND
DL-Menthol	0.0084	ND	ND
Borneol	0.0084	ND	ND
Terpineol	0.0084	ND	ND
Nerol	0.0084	ND	ND
r-Pulegone	0.0084	ND	ND
Geranyl Acetate	0.0084	ND	ND
a-Cedrene	0.0084	ND	ND
b-Caryophyllene	0.0084	ND	ND
a-Humulene	0.0084	ND	ND
Valencene	0.0084	ND	ND
cis-Nerolidol	0.0084	ND	ND
trans-Nerolidol	0.0084	ND	ND
Guaiol	0.0084	ND	ND
Caryophyllene Oxide	0.0084	ND	ND
Cedrol	0.0084	ND	ND
a-Bisabolol	0.0084	ND	ND
m-Cymene	0.0084	ND	ND
trans-b-Ocimene	0.0084	ND	ND
cis-b-Ocimene	0.0084	ND	ND
o-Cymene	0.0084	ND	ND
a-Thujone	0.0084	ND	ND

ACCREDITED
CERT #7425.03

* Note: Analytes tested at ACT Laboratories of FL (ISO/IEC 17025:2017 #7425.05): 1,1-Dichloroethene; Captan, Chlordane, Chlormequat Chloride, Coumaphos, Dimethomorph, Mevinphos, PCNB, Piperonyl Butoxide, Spinetoram, Fenhaxamide.

Brad Cavinder
Laboratory Manager

♦ indicates ISO/IEC 17025:2017 accreditation is pending. This sample has been tested by ACT Laboratories using valid, ISO/IEC 17025:2017 accredited testing methodologies and a quality system as required by state law. Values reported relate only to the sample tested. Results are deemed acceptable if they fall within the specified acceptance criteria as defined by the applicable regulatory limit or client specification, with no measurement uncertainty applied. ACT Laboratories makes no claims as to the efficacy, safety or other risks associated with any detected or non-detected levels of any compounds reported herein. This Certificate shall not be reproduced except in full, without the written approval of ACT Laboratories. The authenticity of this document is only guaranteed if issued from an @actlab.com email address.

Certificate of Analysis

Amended

Compliance Adult Use

SC-000018,AU-SC-000106
ACT Laboratories (MI)
617 E. Hazel, Lansing, Michigan
5178621857
blakefeighner@actlab.com

Nowadays
Order No.: OMIELBN0903-0008682
150 Progress
California, 92618
info@trynowadays.com
7143674386

Sample: SMIELBN0915-IBEV-0006157
Unit Weight: 45.0000g
Sample Size: 2.0000 Units
Date Sampled: 09/03/2025 13:25
Sample Received: 09/03/2025 13:33
Report Created: 09/30/2025 12:46
Sampling SOP 204

CR-NOW750-05-32 - Nowadays - Original - 5mg
Ingestible, Beverage
Batch ID: CR-NOW750-05-32, Lot ID: See Page 9 for more details

Analyte	LOQ (%)	mg/g	%
Menthone	0.0084	ND	ND
Terpinen-4-ol	0.0084	ND	ND
Isomenthone	0.0084	ND	ND
Safranal	0.0084	ND	ND
Citronellol	0.0084	ND	ND
Octyl Acetate	0.0084	ND	ND
Verbenone	0.0084	ND	ND
d-Valerolactam	0.0084	ND	ND
cis-Citral	0.0084	ND	ND
Thymol	0.0084	ND	ND
s-Carvone	0.0084	ND	ND
Carvacrol	0.0167	ND	ND
trans-b-Farnesene	0.0084	ND	ND
Farnesol	0.0084	ND	ND
Phytane	0.0084	ND	ND
Nootkatone	0.0084	ND	ND
Total Terpenes		ND	ND

Report notes:
LOQ = Limit of Quantitation, ND= Non-Detect; The reported result is based on a sample weight with the applicable moisture content for that sample; Unless otherwise stated all quality control samples performed within specification established by the Laboratory. Terpene analysis result is an estimate as the sample response is typically outside the standard curve linear response.

ACCREDITED
CERT #7425.03

* Note: Analytes tested at ACT Laboratories of FL (ISO/IEC 17025:2017 #7425.05): 1,1-Dichloroethene; Captan, Chlordane, Chlormequat Chloride, Coumaphos, Dimethomorph, Mevinphos, PCNB, Piperonyl Butoxide, Spinetoram, Fenhaxamide.

Brad Cavinder
Laboratory Manager

♦ indicates ISO/IEC 17025:2017 accreditation is pending. This sample has been tested by ACT Laboratories using valid, ISO/IEC 17025:2017 accredited testing methodologies and a quality system as required by state law. Values reported relate only to the sample tested. Results are deemed acceptable if they fall within the specified acceptance criteria as defined by the applicable regulatory limit or client specification, with no measurement uncertainty applied. ACT Laboratories makes no claims as to the efficacy, safety or other risks associated with any detected or non-detected levels of any compounds reported herein. This Certificate shall not be reproduced except in full, without the written approval of ACT Laboratories. The authenticity of this document is only guaranteed if issued from an @actlab.com email address.

Certificate of Analysis

Amended

Compliance Adult Use

SC-000018,AU-SC-000106
ACT Laboratories (MI)
617 E. Hazel, Lansing, Michigan
5178621857
blakefeighner@actlab.com

Nowadays
Order No.: OMIELBN0903-0008682
150 Progress
California, 92618
info@trynowadays.com
7143674386

Sample: SMIELBN0915-IBEV-0006157
Unit Weight: 45.0000g
Sample Size: 2.0000 Units
Date Sampled: 09/03/2025 13:25
Sample Received: 09/03/2025 13:33
Report Created: 09/30/2025 12:46
Sampling SOP 204

CR-NOW750-05-32 - Nowadays - Original - 5mg
Ingestible, Beverage
Batch ID: CR-NOW750-05-32, Lot ID: See Page 9 for more details

Phytol

Tested

SOP: SOP 806-MI
Instrument: LCMS 1260-6470 Agilent - 2300

Analyte	LOQ (ppm)	%
Phytol	1,025.78	ND

Residual Solvents

Pass

SOP: SOP 601-MI
Instrument: HS-GC-MS 7697-7890-5977 Agilent - 2651

Analyte	LOQ (ppm)	Limit (ppm)	ppm	%	Status
1,2-Dichloroethane	1	5	ND	Passed	
Acetone	7	5,000	ND	Passed	
Acetonitrile	6	410	ND	Passed	
Benzene	1	2	ND	Passed	
n-Butane	70	5,000	ND	Passed	
Chloroform	1	60	ND	Passed	
Ethanol	8	5,000	ND	Passed	
Ethyl Acetate	10	5,000	ND	Passed	
Ethyl Ether	12	5,000	ND	Passed	
Ethylene Oxide	1	50	ND	Passed	
Heptane	21	5,000	ND	Passed	
n-Hexane	34	290	ND	Passed	
Isopropyl Alcohol	2	5,000	ND	Passed	
Isobutane	70	5,000	ND	Passed	
Methanol	14	3,000	ND	Passed	
Methylene Chloride	1	600	ND	Passed	

Analyte	LOQ (ppm)	Limit (ppm)	ppm	%	Status
Propane	26	5,000	ND	Passed	
Toluene	12	890	ND	Passed	
Trichloroethylene	1	80	ND	Passed	
Xylenes	31	2,170	ND	Passed	
Neopentane	34	5,000	ND	Passed	
Isopentane	70	5,000	ND	Passed	
n-Pentane	39	5,000	ND	Passed	
2,2-Dimethyl-Butane	39	290	ND	Passed	
2,3-Dimethyl-Butane	39	290	ND	Passed	
2-Methyl-Pentane	701	290	ND	Passed	
3-Methyl-Pentane	39	290	ND	Passed	
Total Butanes	7	5,000	ND	Passed	
Total Hexanes	3	290	ND	Passed	
Total Pentanes	4	5,000	ND	Passed	
1,1-Dichloroethene	2	8	ND	Passed	

Report notes:
ND= Non-Detect

ACCREDITED
CERT #7425.03

* Note: Analytes tested at ACT Laboratories of FL (ISO/IEC 17025:2017 #7425.05): 1,1-Dichloroethene; Captan, Chlordane, Chlormequat Chloride, Coumaphos, Dimethomorph, Mevinphos, PCNB, Piperonyl Butoxide, Spinetoram, Fenhaxamide.

Brad Cavinder
Laboratory Manager

♦ indicates ISO/IEC 17025:2017 accreditation is pending. This sample has been tested by ACT Laboratories using valid, ISO/IEC 17025:2017 accredited testing methodologies and a quality system as required by state law. Values reported relate only to the sample tested. Results are deemed acceptable if they fall within the specified acceptance criteria as defined by the applicable regulatory limit or client specification, with no measurement uncertainty applied. ACT Laboratories makes no claims as to the efficacy, safety or other risks associated with any detected or non-detected levels of any compounds reported herein. This Certificate shall not be reproduced except in full, without the written approval of ACT Laboratories. The authenticity of this document is only guaranteed if issued from an @actlab.com email address.

Certificate of Analysis

Amended

Compliance Adult Use

SC-000018,AU-SC-000106
ACT Laboratories (MI)
617 E. Hazel, Lansing, Michigan
5178621857
blakefeighner@actlab.com

Nowadays
Order No.: OMIELBN0903-0008682
150 Progress
California, 92618
info@trynowadays.com
7143674386

Sample: SMIELBN0915-IBEV-0006157
Unit Weight: 45.0000g
Sample Size: 2.0000 Units
Date Sampled: 09/03/2025 13:25
Sample Received: 09/03/2025 13:33
Report Created: 09/30/2025 12:46
Sampling SOP 204

CR-NOW750-05-32 - Nowadays - Original - 5mg
Ingestible, Beverage
Batch ID: CR-NOW750-05-32, Lot ID: See Page 9 for more details

Mycotoxins

Pass

SOP: SOP 808-MI
Instrument: LCMS 1260-6470 Agilent - 2300

Analyte	LOQ (ppb)	Limit (ppb)	ppb	Status
B1	7.93	20.00	ND	Passed
B2	10.21	20.00	ND	Passed
G1	9.12	20.00	ND	Passed
G2	12.52	20.00	ND	Passed
Ochratoxin A	10.73	20.00	ND	Passed
Total Aflatoxins			ND	Tested

Report notes:
LOQ = Limit of Quantitation, ND= Non-Detect; The reported result is based on a sample weight with the applicable moisture content for that sample; Unless otherwise stated all quality control samples performed within specification established by the Laboratory.

Heavy Metals

Pass

SOP: SOP 901-MI, SOP 914
Instrument: ICPMS 7850 Agilent - 1933

Analyte	LOQ (ppm)	Limit (ppm)	ppm	Status
Arsenic	0.073	1.500	ND	Passed
Cadmium	0.073	0.500	ND	Passed
Chromium	0.073	2.000	ND	Passed
Copper	0.725		ND	Tested
Mercury	0.073	3.000	ND	Passed
Nickel	0.218		ND	Tested
Lead	0.218	0.500	ND	Passed

Report notes:
LOQ = Limit of Quantitation, ND= Non-Detect; The reported result is based on a sample weight with the applicable moisture content for that sample; Unless otherwise stated all quality control samples performed within specification established by the Laboratory.

ACCREDITED
CERT #7425.03

* Note: Analytes tested at ACT Laboratories of FL (ISO/IEC 17025:2017 #7425.05): 1,1-Dichloroethene; Captan, Chlordane, Chlormequat Chloride, Coumaphos, Dimethomorph, Mevinphos, PCNB, Piperonyl Butoxide, Spinetoram, Fenhaxamide.

Brad Cavinder
Laboratory Manager

♦ indicates ISO/IEC 17025:2017 accreditation is pending. This sample has been tested by ACT Laboratories using valid, ISO/IEC 17025:2017 accredited testing methodologies and a quality system as required by state law. Values reported relate only to the sample tested. Results are deemed acceptable if they fall within the specified acceptance criteria as defined by the applicable regulatory limit or client specification, with no measurement uncertainty applied. ACT Laboratories makes no claims as to the efficacy, safety or other risks associated with any detected or non-detected levels of any compounds reported herein. This Certificate shall not be reproduced except in full, without the written approval of ACT Laboratories. The authenticity of this document is only guaranteed if issued from an @actlab.com email address.

Certificate of Analysis

Amended

Compliance Adult Use

SC-000018,AU-SC-000106
ACT Laboratories (MI)
617 E. Hazel, Lansing, Michigan
5178621857
blakefeighner@actlab.com

Nowadays
Order No.: OMIELBN0903-0008682
150 Progress
California, 92618
info@trynowadays.com
7143674386

Sample: SMIELBN0915-IBEV-0006157
Unit Weight: 45.0000g
Sample Size: 2.0000 Units
Date Sampled: 09/03/2025 13:25
Sample Received: 09/03/2025 13:33
Report Created: 09/30/2025 12:46
Sampling SOP 204

CR-NOW750-05-32 - Nowadays - Original - 5mg
Ingestible, Beverage
Batch ID: CR-NOW750-05-32, Lot ID: See Page 9 for more details

Pesticides

Pass

SOP: SOP 804-MI, SOP 814-MI
Instrument: LCMS 1260-6470 Agilent - 2300 / GCMS 7890-7010 Agilent - 1254

Analyte	LOQ (ppm)	Limit (ppm)	ppm	Status
Abamectin	0.25	0.5	ND	Passed
Acephate	0.2	0.4	ND	Passed
Acequinocyl	1	2	ND	Passed
Acetamiprid	0.1	0.2	ND	Passed
Aldicarb	0.2	0.4	ND	Passed
Azoxystrobin	0.1	0.2	ND	Passed
Bifenazate	0.1	0.2	ND	Passed
Bifenthrin	0.1	0.2	ND	Passed
Boscalid	0.2	0.4	ND	Passed
Carbaryl	0.1	0.2	ND	Passed
Carbofuran	0.1	0.2	ND	Passed
Chlorantraniliprole	0.1	0.2	ND	Passed
Chlorpyrifos	0.1	0.2	ND	Passed
Clofentezine	0.1	0.2	ND	Passed
Coumaphos	0.05	0.1	ND	Passed
Cyfluthrin	0.5	1	ND	Passed
Cypermethrin	0.5	1	ND	Passed
Daminozide	0.5	1	ND	Passed
Diazinon	0.1	0.2	ND	Passed
Dichlorvos	0.5	1	ND	Passed
Dimethoate	0.1	0.2	ND	Passed
Dimethomorph	0.01	3	ND	Passed
Ethoprophos	0.1	0.2	ND	Passed
Etofenprox	0.2	0.4	ND	Passed
Etoxazole	0.1	0.2	ND	Passed
Fenhexamid	0.047	3	ND	Passed
Fenoxycarb	0.1	0.2	ND	Passed
Fenpyroximate	0.2	0.4	ND	Passed
Fipronil	0.2	0.4	ND	Passed
Flonicamid	0.5	1	ND	Passed
Fludioxonil	0.2	0.4	ND	Passed
Hexythiazox	0.5	1	ND	Passed
Imazalil	0.1	0.2	ND	Passed
Imidacloprid	0.2	0.4	ND	Passed
Kresoxim Methyl	0.2	0.4	ND	Passed
Malathion	0.1	0.2	ND	Passed
Mevinphos	0.021	0.1	ND	Passed
Metalaxyl	0.1	0.2	ND	Passed
Methiocarb	0.1	0.2	ND	Passed
Methomyl	0.2	0.4	ND	Passed

ACCREDITED
CERT #7425.03

* Note: Analytes tested at ACT Laboratories of FL (ISO/IEC 17025:2017 #7425.05): 1,1-Dichloroethene; Captan, Chlordane, Chlormequat Chloride, Coumaphos, Dimethomorph, Mevinphos, PCNB, Piperonyl Butoxide, Spinetoram, Fenhaxamide.

Brad Cavinder
Laboratory Manager

♦ indicates ISO/IEC 17025:2017 accreditation is pending. This sample has been tested by ACT Laboratories using valid, ISO/IEC 17025:2017 accredited testing methodologies and a quality system as required by state law. Values reported relate only to the sample tested. Results are deemed acceptable if they fall within the specified acceptance criteria as defined by the applicable regulatory limit or client specification, with no measurement uncertainty applied. ACT Laboratories makes no claims as to the efficacy, safety or other risks associated with any detected or non-detected levels of any compounds reported herein. This Certificate shall not be reproduced except in full, without the written approval of ACT Laboratories. The authenticity of this document is only guaranteed if issued from an @actlab.com email address.

Certificate of Analysis

Amended

Compliance | Adult Use

SC-000018,AU-SC-000106
ACT Laboratories (MI)
617 E. Hazel, Lansing, Michigan
5178621857
blakefeighner@actlab.com

Nowadays
Order No.: OMIELBN0903-0008682
150 Progress
California, 92618
info@trynowadays.com
7143674386

Sample: SMIELBN0915-IBEV-0006157
Unit Weight: 45.0000g
Sample Size: 2.0000 Units
Date Sampled: 09/03/2025 13:25
Sample Received: 09/03/2025 13:33
Report Created: 09/30/2025 12:46
Sampling SOP 204

CR-NOW750-05-32 - Nowadays - Original - 5mg
Ingestible, Beverage
Batch ID: CR-NOW750-05-32, Lot ID: See Page 9 for more details

Analyte	LOQ (ppm)	Limit (ppm)	ppm	Status
MGK-264	0.1	0.2	ND	Passed
Myclobutanil	0.1	0.2	ND	Passed
Naled	0.25	0.5	ND	Passed
Oxamyl	0.5	1	ND	Passed
Paclobutrazol	0.2	0.4	ND	Passed
Permethrin	0.1	0.2	ND	Passed
Phosmet	0.1	0.2	ND	Passed
Piperonyl Butoxide	0.01	3	ND	Passed
Prallethrin	0.1	0.2	ND	Passed
Propiconazole	0.2	0.4	ND	Passed
Propoxur	0.1	0.2	ND	Passed
Pyrethrins	0.5	1	ND	Passed
Pyridaben	0.1	0.2	ND	Passed
Spinetoram	0.017	3	ND	Passed
Spinosyn AD	0.1	0.2	ND	Passed
Spiromesifen	0.1	0.2	ND	Passed
Spirotetramat	0.1	0.2	ND	Passed
Spiroxamine	0.2	0.4	ND	Passed
Tebuconazole	0.2	0.4	ND	Passed
Thiacloprid	0.1	0.2	ND	Passed
Thiamethoxam	0.1	0.2	ND	Passed
Trifloxystrobin	0.1	0.2	ND	Passed
Captan	0.1	3	ND	Passed
Methyl Parathion	0.1	0.2	ND	Passed
Chlordane	0.02	0.1	0.0000	Passed
Chlorfenapyr	0.5	1	ND	Passed
Chlormequat Chloride	0.01	3	ND	Passed
Pentachloronitrobenzene	0.05	0.2	0.0000	Passed

Report notes:
LOQ = Limit of Quantitation, ND= Non-Detect; The reported result is based on a sample weight with the applicable moisture content for that sample; Unless otherwise stated all quality control samples performed within specification established by the Laboratory.

ACCREDITED
CERT #7425.03

* Note: Analytes tested at ACT Laboratories of FL (ISO/IEC 17025:2017 #7425.05): 1,1-Dichloroethene; Captan, Chlordane, Chlormequat Chloride, Coumaphos, Dimethomorph, Mevinphos, PCNB, Piperonyl Butoxide, Spinetoram, Fenhaxamide.

Brad Cavinder
Laboratory Manager

♦ indicates ISO/IEC 17025:2017 accreditation is pending. This sample has been tested by ACT Laboratories using valid, ISO/IEC 17025:2017 accredited testing methodologies and a quality system as required by state law. Values reported relate only to the sample tested. Results are deemed acceptable if they fall within the specified acceptance criteria as defined by the applicable regulatory limit or client specification, with no measurement uncertainty applied. ACT Laboratories makes no claims as to the efficacy, safety or other risks associated with any detected or non-detected levels of any compounds reported herein. This Certificate shall not be reproduced except in full, without the written approval of ACT Laboratories. The authenticity of this document is only guaranteed if issued from an @actlab.com email address.

Certificate of Analysis

Amended

Compliance Adult Use

SC-000018,AU-SC-000106
ACT Laboratories (MI)
617 E. Hazel, Lansing, Michigan
5178621857
blakefeighner@actlab.com

Nowadays
Order No.: OMIELBN0903-0008682
150 Progress
California, 92618
info@trynowadays.com
7143674386

Sample: SMIELBN0915-IBEV-0006157
Unit Weight: 45.0000g
Sample Size: 2.0000 Units
Date Sampled: 09/03/2025 13:25
Sample Received: 09/03/2025 13:33
Report Created: 09/30/2025 12:46
Sampling SOP 204

CR-NOW750-05-32 - Nowadays - Original - 5mg
Ingestible, Beverage
Batch ID: CR-NOW750-05-32, Lot ID: See Page 9 for more details

Miscellaneous

Tested

SOP: SOP 104-MI
Date/Time Tested: 09/08/2025 09:56

Analyte	Result
pH	**2.84 ph**

Assay Note:
Processor: Ellison Brewery 4903 Dawn Ave. East Lansing, MI 48823 Hemp Lic: HPHL-002445 MI Food Lic: FFP-118497 Facility meets sanitization requirements as documented by the regulatory entity.

ACCREDITED
CERT #7425.03

* Note: Analytes tested at ACT Laboratories of FL (ISO/IEC 17025:2017 #7425.05): 1,1-Dichloroethene; Captan, Chlordane, Chlormequat Chloride, Coumaphos, Dimethomorph, Mevinphos, PCNB, Piperonyl Butoxide, Spinetoram, Fenhaxamide.

Brad Cavinder
Laboratory Manager

♦ indicates ISO/IEC 17025:2017 accreditation is pending. This sample has been tested by ACT Laboratories using valid, ISO/IEC 17025:2017 accredited testing methodologies and a quality system as required by state law. Values reported relate only to the sample tested. Results are deemed acceptable if they fall within the specified acceptance criteria as defined by the applicable regulatory limit or client specification, with no measurement uncertainty applied. ACT Laboratories makes no claims as to the efficacy, safety or other risks associated with any detected or non-detected levels of any compounds reported herein. This Certificate shall not be reproduced except in full, without the written approval of ACT Laboratories. The authenticity of this document is only guaranteed if issued from an @actlab.com email address.